

Clarifications/ Amendments

IVCS for mice (NIT No NIPER/RBL/2020-21/20/273)

S. No.	Queries raised by the prospective bidders	Clarifications/ Amendments in response to the queries
1.	We would request you to kindly change the number of autoclave cycles as 150 – 200 cycles for autoclavable parts, as it is dependent on autoclaving parameters and also would enable all the seller to participated in the tender	Should read the clause as “General conditions: Construction material for autoclavable parts should have high grade quality and should not defade/de-color/leached/change-in-color on autoclave for at-least 150 cycles. Supplier must provide undertaking on letter-head signed by authorized person of the company/supplier/ manufacturer with designation”
2.	Kindly clarify if 5KVA UPS is required and the quantity required as this would affect the overall costing of the system, we would suggest for 2 number of UPS of 800 VA,12 V which is suitable to provide a backup of 30 min for both the IVC systems	UPS (one) with 5KVA capacity (minimum 30 min backup) should be included in the system.

Pre-bid meeting: 30/08/2021 at 15.30 pm (online)